

## NOTICE INVITING TENDER

Tender No. DPMSU-KNR/404/AC/2020 dated 04-06-2020

The District Program Manager, National Health Mission, Civil Station, Kannur - 670002 (Phone No. 0497 2709920, E-mail : dpmknr@gmail.com) invites Online Bids from reputed manufactures/suppliers through e-tendering for the supply and installation of OT Light Double Dome as details furnished below.

Name of work	<b>Supply and installation of Two OT Light Double Dome at TH Payyanur and TH Kuthuparamba ( One each)</b>
Cost of Tender Form	<b>Rs. 2020/- (including GST)</b>
Publishing Date and time	<b>05/06/2020, 11.00 AM</b>
Document Sale/download Start Date	<b>05/06/2020, 11.00 AM</b>
Bid Submission Start Date	<b>05/06/2020, 11.00 AM</b>
Bid Submission End date	<b>12/06/2020, 11.00 AM</b>
Bid Opening Date	<b>15/06/2020, 2.30 PM</b>

### Scope of work

Supply and installation of Two OT Light Double Dome at TH Payyanur and TH Kuthuparamba (One each) as per detailed specification attached herewith.

### TERMS AND CONDITIONS

1. The bidders should have the necessary portal enrollment (with their own Digital Signature Certificate)
2. Tender can be submitted online on the website **etenders.kerala.gov.in**.
3. A non-refundable tender document fee of **Rs 2020/-** shall be remitted to the account number in the remittance form provided by the e-procurement system for this particular tender. Only **RTGS/NEFT/SBI Internet Banking** remittance is allowed.
4. An EARNEST MONEY DEPOSIT (EMD) for an amount of **Rs 9000/-(Rupees Nine Thousand only)** shall be remitted online.
5. The offers shall be submitted in Two Bid System i.e. separate Technical & Financial bid required.
6. All the tender documents are to be submitted online only and in the designated covers on the above website and no manual submission shall be entertained.
7. All communications regarding the tender including queries if any shall be done online within the e-procurement system at website **http:// etenders.kerala. gov.in**
8. The bid shall be opened online at the office of the District Program Manager, National Health Mission, Civil Station, Kannur - 670002 on the date and time mentioned above in the presence of bidders / their representatives who are present.
9. The successful bidder shall execute an agreement in the form prescribed by Govt. of Kerala for the due fulfillment of the contract within the period to be specified in the purchase order and shall furnish Security Deposit equal to 5% of the contract value before signing the agreement.
10. All orders and directions issued by the Govt. in connection with inviting tenders, E-Tenders, Store Purchase Rules etc. are also applicable to this E-Tender.

11. More details about the e-tendering procedure are available from Kerala State IT Mission on all working days from 10.00 am to 5.30 pm (Ph: 0471-2577088, 2577188, 2577388 Ph: 0484 2336006, 2332262 or email: etendershelp@kerala.gov.in
12. Detailed specification of the equipment is attached herewith.

Dr. Latheesh K V

Place: Kannur

District Programme Manager

Date: 04-06-2020

National Health Mission, Kannur

## Approval Valid

Digitally Approved By

Dr. Latheesh K.V

Date: 04.06.2020

Reason: Approved

The document is digitally approved. Hence signature is not needed.

## DETAILED SPECIFICATIONS

### Equipment: OT Light shadows less ceiling mount double dome

1. Double dome ceiling type Operation Theater light shadow less LED Type
2. Light output should be minimum 1,00,000 lux at 1 meter distance for each dome
3. Diameter of the dome should be less than 650mm
4. Minimum light field size should be 200mm
5. Power supply to the lamp should be from a constant voltage source/SMPS
6. Color temperature should be between 4000Esk to 5000 Esk
7. Should have short circuit protection for the low voltage power supply
8. Should have heat reduction filters
9. Should have field focus control through sterile handle. Additional sterile handles two numbers should be supplied.
10. All LEDs should be replaced free of cost during warranty/CAMC
11. Minimum spring arm stroke of 500mm
12. Minimum four step intensity control
13. Should have minimum action radius of 1500mm
14. Input supply -230Vac.50Hz
15. Should have safely certificate from a competent authority CE issued by notified body registered in the European commission/FDA(US)/BIS certificate. Copy of the certificate/test report shall be produced along with the technical bid.