



SHORT TENDER NOTICE

Tender Number: NPCB.68/2021 Dated: 09/01/2021

Sealed Tenders are invited from the reputed firms who are willing to supply and install the following item on or before 18/01/2021 at Idukki District.

1.Ophthalmic Equipments (various types)

Last date of Submission of tenders - 18/01/2021, 12.00 PM

Opening Date of the Tender - 18/01/2021 , 2.00 PM .

EMD: - Rs. 2530/- (D.D), Cost of Tender Form :- Rs 600/- (D.D)

(Address of D.D , should be drawn in favour of "DIST HLT & FMLY WELFARE SOCIETY DBCS IDUKKI"- payable at KATTAPPANA). For more details of Tender specifications visit Arogyakeralam website : www.arogyakeralam.gov.in



*Dist. Programme Manager
Blindness Control Society, Idukki*

District programme Officer,

NPCB, Idukki

PH : 04862 232221

Items with Technical specifications

Sl.no	Item	Total Quantity	Location (One Quantity each for mentioned location)
1	Streak Retinoscope Rechargeable type	5 no.	FHC Kudayathoor
2	Trial Frame Adult	5 no.	
3	Trial Frame Child	5 no.	
4	Illuminated Vision Drum Rotating Type	5 no.	PHC Elappara
5	Binomag/ Magnifying Loupe	5 no.	PHC Bisonvally
6	Epilation Forceps	5 no.	
7	Foreign Body Spud	5 no.	PHC Santhanpara
8	LE Torch Rechargeable with 2 Batteries	5 no.	
9	Vision charts 6 Meter Number	5 no.	CHC Kanjikuzhy
10	Wooden trial set	5 no.	
11	Mini trial sets (Pair type)	5 no.	
12	Ishihara Colour Vision Chart	5 no.	

Technical Specifications – Streak Retinoscope

1. Should have an external focusing sleeve which is easy to grip and manipulate.
2. Should have crossed –linear polarizing filter.
3. Should allow one-hand operation for streak focus and 360 streak rotation.
4. Should be interchangeable to plane mirror and concave mirror mode by sleeve movement.
5. Should have holoign / LED light source.
6. Should have 100% dust proof housing and multi-coated optics.
7. Should have brow rest for spectacle wearer.
8. Should be rechargeable battery with Charger.
9. Should have a carrying case.
10. If halogen lamp is used ,then the following additional accessories Should be supplied.